

Compliance Summary

regarding
Electrical Safety and Electromagnetic Compatibility and Environmental Conditions Requirements



CEcert GmbH, Alter Holzhafen 19a, 23966 Wismar (Germany)

Test Laboratory

No.: 619.306.1

License Holder:	Dräger Safety AG & Co. KGaA	
	Revalstraße 1	
	23560 Lübeck, Germany	
Product identification:	Breath alcohol test device	
Type/Model/Parameters:	Alcotest® 6000	
S/N:	ARHC-0014; ARHL-0003; ARHC-0015; ARHL-0008; ARJB-0004; ARJB-0003; ARJB-0002; ARJA-0010ARNC-0006; ARNC-0001	
Applied standards:	EN 15964:2011 Breath alcohol test devices other than single use devices – Requirements and test methods	
Test report(s):	415.361.1, 415.361.10D, 415.361.11 419.306.1	Dates tests performed: 2015-11-26 – 2018-11-28 2020-04-07 – 2020-04-30
Test result:	PASS	

This test summary is relevant exclusively to the item(s) submitted for testing. This test summary is intended as cover sheet for the test report referenced. For details, see the test report specified in this summary and the safety instructions specified in the technical documentation to the product.

Approved by

M.Eng. Philipp Meiske

Wismar, Thursday, 28 May 2020

see 'Attachment 1' for more details

General information:

Tests performed / Restrictions: **all tests in accordance to the applied standard were conducted**
Sampling: **samples used for tests allocated by manufacturer / applicant**

Classification and acceptance:

Design Particulars: **hand-held device**
Measurement range: **0 to 2.5 mg/l**
Result indication: **quantitative (digital)**
Ratings: **1x Lithium battery 3 V, 1400 mAh**

Temperature range (operation): **-5 °C to +50 °C**
Temperature range (storage): **-20 °C to +70 °C**
Calibration interval: **36 months**

End of Listing!