



DECLARATION OF CONFORMITY

DoC No.: BTTF-010 1.0

Manufacturer: Hangzhou Bioer Technology Co., Ltd.
Address: 1192 Bin An Rd, Hi-tech (Binjiang) District, Hangzhou,
310053, P. R. China

Whose **Single** CMC MEDICAL DEVICES & DRUGS SL
Authorized C/Horacio Lengo N° 18 CP 29006, Málaga-Spain

EU-Representative: Tel: +34951214054
Fax: +34952330100

Product Name: SARS-CoV-2 /Influenza A Virus / Influenza B Virus Nucleic Acid
Detection Kit (Fluorescence RT-PCR)

Cat. Number: BSJ17S1 BSJ17M1

Classification: Others of ANNEX II of IVDD

Conformity Assessment Route: Annex III

Applicable Standards: EN ISO 13485:2016 EN ISO 15223-1:2016
EN ISO 14971:2019 EN 13641: 2002
EN ISO 18113-1:2011 EN 13612: 2002
EN ISO 18113-2:2011 EN ISO 23640:2015

We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

General applicable directives:

In Vitro Diagnostic Medical Devices DIRECTIVE 98/79/EC

Signature:

Name:

Title:

Position:

Date:

General manager
Hangzhou, China
2020-12-08

