

CardioSoft™ Ambulatory Blood Pressure

A broader perspective on cardiac patient management

Ambulatory blood pressure (ABP) readings over time provide critical data. Only an ABP device that is simple to set-up, comfortable to wear, and quick to report will help support high patient compliance and accurate diagnosis.



TONOPORT™ VI meets every requirement

Simple, flexible programming. The TONOPORT VI APB module is simple to set up and program to ensure accurate, validated² ABP readings and analysis. BP readings can be set to exact intervals or captured randomly, with day and night programming options.

High-comfort cuff. Monitoring is quiet, comfortable and quick with TONOPORT VI. The innovative inflation measurement method, lightweight design, and low-noise pump enhance patient comfort – helping to increase acceptance of extended monitoring.

- 50% faster inflation with lower maximum pressure
- Quiet pump operation at 40dB – equivalent to a hushed library³

Quick-view trend summaries and reporting. Recorded data is easily downloaded and reported via the CardioSoft Cardiac Testing System. The physician sees a comprehensive data set, including up to 72-hour blood pressure trends, averages and statistics for day and night summaries, presented in text and graphics. Reports can be easily exported to EMRs, PACS and MUSE™ systems.

Simple. Comfortable. Quick.

High blood pressure is a major risk factor for

CORONARY

HEART DISEASE

as well as ischemic and hemorrhagic stroke



Worldwide, hypertension is estimated to cause

7.5 MILLION DEATHS¹



Computer specifications

Microprocessor	Minimum Pentium® 4 class processor with 2 GHz	Acquisition period	Up to 400 measurements or 3 days
RAM	Minimum 2 GB	Interfaces	USB (1.1 or 2.0), RS 232 (9.600Bd / 8N1)
Hard drive	Minimum 80 GB and 4 GB of free space if used as a standalone system	Battery	2AA size rechargeable NiMH batteries, 1.2 V, >1500 mAh or 2 AA size high-current capable alkaline batteries
SW installation	DVD-ROM drive or USB	Battery charge time	2 to 3 hours
Pointer	Mouse	Battery charger	Protection class II, IP20 Primary 100 to 240 VAC 50/60 Hz, 0.5 A
Display resolution	Minimum: 1280 x 768 Maximum: 3840 x 2160	Maximum cuff pressure	300 mmHg
Interfaces	Minimum: 2 USB ports (1.1, 2.0, or 3.0) for each device using this type of interface, CD-RW, SD card, network interface card (recommended), Serial RS232 for each device using this interface type	Measurement Method	Oscillometric, selectable measurement method: deflation measurement method or inflation measurement method
Operating system	Windows® 10 Enterprise (64 bit) Windows 10 Professional (64 bit) Windows 8.1 Enterprise (64 bit) Window 8.1 Pro (64 bit) Windows 7 Professional (64 bit) with SP1	Audible signal	Configurable audio beep before every measurement
Printer	Equivalent to HP® P3015dn (Customer Supplied)	Inflation noise	40 dB
Additional software for export (optional, Customer Supplied) functionality	Microsoft® Word and Excel®	Dimensions and weight of recorder	Height: 27 mm Width: 73 mm Depth: 108 mm Weight: <190 g, including batteries
Networking LAN	Wired and Wireless: 802.11 G (optional) TCP/IP interface	Protection Class	IP22: TONOPORT VI in Wearable Pouch
		Validations	BHS, ESH, ANSI/AAMI SP10, recommended by dabl Educational Trust

Ambulatory BP Specification

Measuring range	Systolic pressure: 60–260 mmHg (8.0–34.6 kPa) Diastolic pressure: 40–220 mmHg (5.3–29.3 kPa) Mean pressure: 50–250 mmHg (6.7–33.3 kPa) Heart rate (HR): 35–240 beats per minute	Transport and storage	Temperature: 5 to 40° C Relative humidity: 15-93%, non-condensing Atmospheric pressure: 700-1060 hPa altitude (relative to sea level) -400 to 2800 meters Temperature: -25 to 70° C Relative humidity: 10-93%, non-condensing Atmospheric pressure: 500-1060 hPa altitude (relative to sea level) -400 to 4500 meters
-----------------	--	-----------------------	--

1 Raised blood pressure. Global Health Observatory data. World Health Organization. http://www.who.int/gho/ncd/risk_factors/blood_pressure_prevalence_text/en/

2 TONOPORT VI ABP device has BIHS, ESH, ANSI/AAMI SP10 validation

3 Noise sources and their effects. Purdue University Chemistry Department. <https://www.chem.purdue.edu/chemsafety/Training/PPETrain/dblevels.htm>

© 2019 General Electric Company – All rights reserved.

GE Healthcare reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your GE Healthcare representative for the most current information. GE, the GE Monogram, CardioSoft, MUSE and TONOPORT are trademarks of General Electric Company. GE Healthcare, a division of General Electric Company, dabi is a trademark of dabl Educational Trust, LTD. HP is a trademark of Hewlett-Packard, Inc. Pentium is a trademark of Intel Corporation. Excel, Microsoft and Windows are trademarks of Microsoft Corporation. GE Medical Systems, Inc., doing business as GE Healthcare.

CardioSoft v7
JB64484XX

Distributed by:

GE Healthcare

Manufactured by:

 PAR Medizintechnik GmbH & Co. KG
Sachsendamm 6
10829 Berlin
GERMANY