Compliance Summary

regarding

Electrical Safety and Electromagnetic Compatibility and Environmental Conditions Requirements

CEcert GmbH, Alter Holzhafen 19a, 23966 Wismar (Germany)

Test Laboratory

No.: 619.306.1

License Holder:	**Dräger Safety AG & Co. KGaA** **Revalstraße 1** **23560 Lübeck, Germany**	
Product identification:	**Breath alcohol test device**	
Type/Model/Parameters:	**Alcotest® 6000**	
S/N:	ARHC-0014; ARHL-0003; ARHC-0015; ARHL-0008; ARJB-0004; ARJB-0003; ARJB-0002; ARJA-0010ARNC-0006; ARNC-0001	
Applied standards:	**EN 15964:2011** Breath alcohol test devices other than single use devices – Requirements and test methods	
		Dates tests performed:
Test report(s):	**415.361.1, 415.361.10D, 415.361.11**	2015-11-26 – 2018-11-28
	419.306.1	2020-04-07 – 2020-04-30
Test result:	**PASS**	

This test summary is relevant exclusively to the item(s) submitted for testing. This test summary is intended as cover sheet for the test report referenced. For details, see the test report specified in this summary and the safety instructions specified in the technical documentation to the product.

Approved by

Wismar, Thursday, 28 May 2020

M.Eng. Philipp Meiske

see ‘Attachment 1’ for more details

© CEcert GmbH; Alter Holzhafen 19a; D-23966 Wismar (Germany)
Managing director: Bernd Schmidt; Phone +49 (0) 38 41 22 42 906, mail: info@cecert.de

Attachment 1 to certificate 619.306.1

General information:

Tests performed / Restrictions:	**all tests in accordance to the applied standard were conducted**
Sampling:	**samples used for tests allocated by manufacturer / applicant**

Classification and acceptance:

Design Particulars:	**hand-held device**
Measurement range:	**0 to 2.5 mg/l**
Result indication:	**quantitative (digital)**
Ratings:	**1x Lithium battery 3 V, 1400 mAh**
Temperature range (operation):	**-5 °C to +50 °C**
Temperature range (storage):	**-20 °C to +70 °C**
Calibration interval:	**36 months**

End of Listing!

© CEcert GmbH; Alter Holzhafen 19a; D-23966 Wismar (Germany)
Managing director: Bernd Schmidt; Phone +49 (0) 38 41 22 42 906, mail: info@cecert.de